



TESTING KITS FOR DRUGS

Procedure Card for IDT0710

BUPHEDRONE IDENTIFICATION

Substance Test



As a precaution, gloves and eye protection should be worn when handling drug substances and using test units. Do not puncture any part of unit. Do not remove probe once a capsule is broken. In the unlikely event of a spill, rinse affected area thoroughly with water. Keep out of reach of children. All results are presumptive. If necessary, further testing by a competent drug testing lab can provide additional identification.

Test Preparation	1	Remove the Smart-Tip™ probe from test unit.	2 Sample suspected substance.	
			Powder Touch Smart-Tip™ probe to sample. 	Tablet Firmly press Smart-Tip™ probe on the tablet while scrubbing in a circular motion until the tip is completely covered with substance.

Test Preparation	3	Lock Smart-Tip™ probe into test unit. Press firmly. 	STOP Do not continue unless probe is locked below the Flex Pins and cannot be removed from the test unit.	Identification	4	With thumb and forefinger, squeeze chamber 1 until capsule breaks.	5	Gently shake test unit for 2-3 seconds.

Results		The appearance of a yellow color indicates the presence of a Cathinone, or an analogue such as Buphedrone , Mephedrone (4-Methylcathinone), Methylcathinone, or N,N-Dimethylcathinone.
		If no yellow color develops, the above listed drugs are not present.
		Proceed to Dilution (steps 6-9), regardless of whether or not a color appeared.

Dilution	6	With thumb and forefinger, squeeze chamber 2 until capsule breaks. 	7	Gently shake test unit for 2-3 seconds. 	Wait 5 minutes. The test window will warm up. This is normal.
----------	---	--	---	---	--

Dilution	8	With thumb and forefinger, squeeze chamber 3 until capsule breaks. 	9	Gently shake test unit for 2-3 seconds. 	The test window will warm up. This is normal.
----------	---	--	---	---	---

	Disposal	After Dilution, the test kit is safe to dispose of in the trash.
--	-----------------	--

