



As a precaution, gloves and eye protection should be worn when handling drug substances and using test units. Do not puncture any part of unit. Do not remove probe once a capsule is broken. In the unlikely event of a spill, rinse affected area thoroughly with water. Keep out of reach of children. All results are presumptive. If necessary, further testing by a competent drug testing lab can provide additional identification.

Test Preparation	1	Remove the Smart-Tip™ probe from test unit.	2		Sample suspected substance.
			Powder	Tablet	
			Touch Smart-Tip™ probe to sample.		Firmly press Smart-Tip™ probe on the tablet while scrubbing in a circular motion until the tip is completely covered with substance.

Test Preparation	3	Lock Smart-Tip™ probe into test unit. Press firmly.	STOP Do not continue unless probe is locked below the Flex Pins and cannot be removed from the test unit.	Identification	4	With thumb and forefinger, squeeze chamber 1 until capsule breaks.	5	Gently shake test unit for 2-3 seconds.

Results		The appearance of a dark blue-green color indicates the presence of Ecstasy/MDMA or an opiate such as Heroin, Morphine or Codeine .
		If no dark blue-green color develops, the above drugs are not present.
		Proceed to Dilution (steps 6-9), regardless of whether or not a color appeared.

Dilution	6	With thumb and forefinger, squeeze chamber 2 until capsule breaks.	7	Gently shake test unit for 2-3 seconds.
				Wait 5 minutes. The test window will warm up. This is normal.

Dilution	8	With thumb and forefinger, squeeze chamber 3 until capsule breaks.	9	Gently shake test unit for 2-3 seconds.
				The test window will warm up. This is normal.

	Disposal	After Dilution, the test kit is safe to dispose of in the trash.
--	-----------------	--